

# CER Endoscope Reprocessing Fundamentals Practice Test Questions and Answers

## 1. What is the first step in endoscope reprocessing?

- A) High-level disinfection
- B) Pre-cleaning and leak testing
- C) Sterilization
- D) Storage preparation

## 2. Which type of water should be used for the final rinse in endoscope reprocessing?

- A) Tap water
- B) Sterile or bacteria-free water
- C) Distilled water
- D) Saline solution

## 3. What is the minimum concentration of glutaraldehyde typically used for high-level disinfection?

- A) 1.0%
- B) 2.0%
- C) 2.4%
- D) 3.0%

## 4. How long should endoscopes be completely dried before storage?

- A) Until visibly dry
- B) For a specified time period per manufacturer guidelines
- C) 5 minutes minimum
- D) Drying is not necessary

Answers: 1-B 2-B 3-B 4-B

For More CER Endoscope Reprocessing Fundamentals Questions and Answers FREE, CER Endoscope Reprocessing Fundamentals Online Prep Training, CER Endoscope Reprocessing Fundamentals Exam, CER Endoscope Reprocessing Fundamentals Study Guide, CER Endoscope Reprocessing Fundamentals Flashcards, CER Endoscope Reprocessing Fundamentals Quizzes visit: